

KAPA Dual-Indexed Adapter Kit Illumina® Platforms

KR1318 – v3.19

This Technical Data Sheet provides product information and guidelines for use of KAPA Dual-Indexed Adapter Kits for Illumina platforms.

This document applies to the KAPA Dual-Indexed Adapter Kit for Illumina platforms (08278555702), and the standalone KAPA Adapter Dilution Buffer (08278539001).

Contents

Product Description	2
Product Applications	2
Product Specifications	2
Shipping and Storage	2
Procedure for handling Dual-Indexed Adapter Plate ..	2
Quality Control	2
Important Parameters	3
Best Practices	3
Compatibility with KAPA Library Preparation Kits ...	3
Additional Notes on Adapter Concentration	5
Adapter Concentration Calculations	5
Post-ligation Processing	5
Index Sequences and Pooling Guidelines	6
Plate Format	6
Restrictions and Liabilities	8
Note to Purchaser: Limited Product Warranty	8
Note to Purchaser: Limited License	8

KAPA/Roche Kit Codes and Components		
KK8722 08278555702*	KAPA Dual-Indexed Adapter Plate (20 µL/well)	96 x 15 µM
	KAPA Adapter Dilution Buffer	25 mL
	KAPA Adapter Sealing Foils (pierceable)	3 foils
KK8721 08278539001	KAPA Adapter Dilution Buffer	25 mL

* KK8722 (08278555702) = KK8720 (08278512001) + KK8721 (08278539001)

Quick Notes

- KAPADual-IndexedAdaptersarefull-lengthadapters used during ligation-based library construction for sequencing on Illumina® instruments.
- KAPA Dual-Indexed Adapter Kits contain 20 µL of each indexed adapter, supplied at a concentration of 15 µM in plate format.
- The number of libraries that can be prepared with each KAPA Dual-Indexed Adapter Kit is dependent on the amount of input DNA, the average fragment size of the input DNA, and the kit used for library construction. Please refer to **Important Parameters** for guidelines on how to use KAPA Dual-Indexed Adapters in combination with different KAPA library preparation kits. With no dilution, 4 libraries can be prepared with each of the 96 dual-indexed adapters, for a total of 384 libraries per kit.
- KAPA Dual-Indexed Adapters are duplexed oligonucleotides and must not be exposed to temperatures above room temperature. Adapters must be diluted in the KAPA Adapter Dilution Buffer provided in the kit to avoid dissociation and ensure optimal performance.
- Employ best laboratory practices to avoid cross contamination of indexed adapters.
- To ensure equal read distributions in multiplexed sequencing applications, libraries must be carefully quantified and/or normalized prior to pooling for capture or cluster generation. qPCR-based quantification with the KAPA Library Quantification Kit constitutes the most accurate and reproducible method for the quantification of sequenceable molecules. This is particularly true for PCR-free workflows.

Product Description

KAPA Dual-Indexed Adapter Kits for Illumina platforms are full-length adapters used during ligation-based library construction for sequencing on Illumina® instruments. Adapters are designed for use with KAPA DNA and RNA library preparation kits. Each plate contains a set of 96 adapters, each with two, 8-nucleotide indexes (barcodes) for multiplexed sequencing applications.

The sequences of the sequencing indexes (barcodes) included in KAPA Dual-Indexed Adapter Kits are given in **Important Parameters: Index Sequences and Pooling Guidelines** (p. 6). The highly dissimilar nature of indexes minimizes the possibility of inaccurate index calling.

KAPA Adapter Dilution Buffer [10 mM Tris-HCl, (pH 8.0 – 8.5), 10 mM NaCl, 1 mM EDTA] is provided with the kit to ensure optimal performance when adapters require further dilution.

Product Applications

KAPA Dual-Indexed Adapters are used to uniquely label sequencing libraries generated from individual biological samples. This allows for the pooling of libraries prior to target capture or cluster generation, to enable multiplexed sequencing—which simplifies sample preparation and reduces the cost of next-generation sequencing for a wide range of applications.

Primary applications for the use of KAPA Dual-Indexed Adapter Kits for Illumina platforms include:

- Human whole genome sequencing, performed with an Illumina HiSeq X instrument
- whole exome or targeted sequencing, using SeqCap EZ or IDT xGen Lockdown Probes or other hybridization capture systems, in combination with the appropriate blockers
- RNA-seq
- ChIP-seq
- other direct sequencing applications, e.g., microbial whole-genome sequencing on compatible platforms.

NOTE: KAPA Dual-Indexed Adapters are not methylated, and can therefore not be used for methyl-seq applications.

Product Specifications

Shipping and Storage

KAPA Dual-Indexed Adapter Kits are shipped on dry ice or ice packs, depending on the destination country. Upon receipt, immediately store the product at -15°C to -25°C in a constant-temperature freezer. Adapters must not be exposed to temperatures above room temperature. When stored under these conditions and handled correctly, the adapters will retain full functionality until the expiry date indicated on the kit label. The KAPA Adapter Dilution Buffer may be stored at 2°C to 8°C for short-term use, but -15°C to -25°C is recommended for long-term storage.

Procedure for handling Dual-Indexed Adapter Plate

- Remove the KAPA Dual-Indexed Adapter plate from its packaging sleeve and thaw at room temperature for 10 min.
- Centrifuge the KAPA Dual-Indexed Adapter plate at 280 x g for 1 min at room temperature.
- Carefully remove the foil cover to avoid cross-contamination of the dual-indexed adapters. Discard the foil cover.
- Remove the desired volume of each dual-indexed adapter by following the guidelines in **Important Parameters**.
- If you are not processing the entire KAPA Dual-Indexed Adapter plate at the same time, apply a new adhesive foil seal (3 are supplied in the KAPA Dual-Indexed Adapter Kit).
 - If required, similar sealing film can be used for plates containing diluted adapters.
- Label the re-sealed Dual-Indexed Adapter plate and store at -15°C to -25°C in a constant-temperature freezer.

Quality Control

KAPA Dual-Indexed Adapters are subject to stringent functional and barcode cross-contamination quality control. KAPA Adapter Dilution Buffer is free of detectable contaminating exo- and endonuclease activities, and meets strict requirements with respect to DNA contamination.

Important Parameters

Best Practices

- KAPA Dual-Indexed Adapters should be used on ice or in cooled reagent blocks, and must not be exposed to conditions above room temperature.
- Use the KAPA Adapter Dilution Buffer [10 mM Tris-HCl, (pH 8.0 – 8.5), 10 mM NaCl, 1 mM EDTA] provided in the kit to dilute KAPA Dual-Indexed Adapters. Adapters diluted in any other buffer or in PCR-grade water may not support optimal library construction efficiency.
- The KAPA Dual-Indexed Adapter plate contains an excess of each adapter, over and above the stated volume of 20 µL. For this reason, and because diluted adapters are less stable, adapter dilutions should not be performed in the plate in which the adapters are supplied. Dilute only the amount of each adapter needed for same-day usage, in a new plate. Long-term storage and multiple cycles of freezing and thawing of diluted adapter stocks are not recommended.
- Employ best laboratory practices to avoid cross-contamination of adapters and/or the dilution buffer.
- Always use plastics that are certified to be free of DNases, RNases, and nucleases. Low DNA- and RNA-binding plastics are highly recommended, especially for low-input DNA and all RNA-Seq library construction applications.

Compatibility with KAPA Library Preparation Kits

KAPA Dual-Indexed Adapter Kits for Illumina platforms are recommended for use with all KAPA library construction kits.

When selecting a KAPA Dual-Indexed Adapter Kit for a specific application, please consider the following:

- Refer to the table corresponding to the KAPA Library Preparation Kit that will be used (p. 4). Recommended adapter stock concentrations for libraries constructed from different inputs and DNA fragment lengths with the listed kits listed are included in Tables 1 – 7.
- Identify the recommended adapter stock concentration for the input and fragment length.
- If the input and specific fragment length are not listed in the table, follow the instructions in **Adapter Concentration Calculations** (p. 5) to calculate the

appropriate adapter stock concentration for the experiment.

- For each batch of libraries to be constructed, prepare an appropriate volume of diluted adapter using the KAPA Adapter Dilution Buffer provided in the kit. Standard protocols call for 5 µL of appropriately diluted adapter stock per library.
 - If an adapter stock concentration >15 µM is required, the volume of water in the ligation reaction may be reduced and the volume of adapter increased to the same extent, up to a total of 10 µL adapter per reaction.
 - Diluted adapters should be freshly prepared and must not be stored for long periods of time or subjected to repeated freezing and thawing.
 - An excess volume of each diluted adapter stock will be required to ensure accurate dispensing. The excess will be larger for automated vs. manual use.
 - All or some of the libraries in the batch may require a different index. Please refer to **Important Parameters: Index Sequences and Pooling Guidelines** (p. 6) for recommendations on multiplexing.
 - The Technical Data Sheet included with the library preparation kit contains specific guidelines for the optimization of adapter concentration when using that particular kit for different applications.

Table 1. Recommended adapter concentrations for KAPA HyperPrep and KAPA HyperPlus Kits¹

Fragmented DNA per 60 µL ER & AT reaction	Adapter stock concentration	Adapter:insert molar ratio	Fragmented DNA per 60 µL ER & AT reaction	Adapter stock concentration	Adapter:insert molar ratio
1 µg	15 µM	10:1	25 ng	7.5 µM	200:1
500 ng	15 µM	20:1	10 ng	3 µM	200:1
250 ng	15 µM	40:1	5 ng	1.5 µM	200:1
100 ng	15 µM	100:1	2.5 ng	750 nM	200:1
50 ng	15 µM	200:1	1 ng	300 nM	200:1

¹Adapter:insert molar ratio calculations are based on a mode DNA fragment length of 200 bp, and will be higher for longer DNA fragments, or slightly lower for DNA fragmented to a mode size <200 bp. The lower adapter:insert molar ratios recommended for inputs >100 ng represent a fair compromise between library construction efficiency and cost; higher library yields will be achieved if a higher adapter concentration is used.

Table 2. Recommended adapter concentrations for KAPA HTP and LTP Library Preparation Kits¹

Insert DNA per 70 µL end repair reaction	Recommended adapter concentration for DNA sheared to an average size of					
	175 bp		350 bp		500 bp	
	Stock	Final	Stock	Final	Stock	Final
1 µg	20 µM	2 µM	10 µM	1 µM	7 µM	0.7 µM
500 ng	10 µM	1 µM	5 µM	500 nM	3.5 µM	350 nM
100 ng	2 µM	200 nM	1 µM	100 nM	700 nM	70 nM
10 ng	200 nM	20 nM	100 nM	10 nM	70 nM	7 nM

¹Adapter concentrations are based on a ~10:1 adapter:insert ratio. For values not listed here, please refer to the calculation in Important Parameters: Adapter Concentration Calculations or visit sequencing.roche.com/support to obtain a calculator designed for this purpose.

Table 3. Recommended adapter concentrations for KAPA RNA HyperPrep Kits

Quantity of starting material	Starting material quality	Adapter stock concentration
1 – 49 ng	Partially degraded or FFPE-derived	1.5 µM
	High-quality	1.5 µM
50 – 100 ng	Partially degraded or FFPE-derived	1.5 µM
	High-quality	15 µM

Table 5. Recommended adapter concentrations for KAPA RNA HyperPrep Kits with RiboErase (HMR)

Quantity of starting material	Starting material quality	Adapter stock concentration
25 – 499 ng	Partially degraded or FFPE-derived	1.5 µM
	High-quality	1.5 µM
500 – 1000 ng	Partially degraded or FFPE-derived	1.5 µM
	High-quality	7 µM

Table 4. Recommended adapter concentrations for KAPA mRNA HyperPrep Kits

Quantity of starting material	Adapter stock concentration
50 – 499 ng	1.5 µM
500 – 1000 ng	7 µM

Table 6. Recommended adapter concentrations for KAPA Stranded RNA-Seq Kits with RiboErase (HMR)

Input RNA	Adapter stock concentration	Final adapter concentration
501 – 1000 ng	280 nM	20 nM
251 – 500 ng	210 nM	15 nM
100 – 250 ng	140 nM	10 nM

Table 7. Recommended adapter concentrations for KAPA Stranded RNA and mRNA-Seq Library Preparation Kits

Input RNA		Adapter stock concentration	Final adapter concentration
RNA workflow	mRNA workflow		
201 – 400 ng	2001 – 4000 ng	1400 nM	20 nM
51 – 200 ng	501 – 2000 ng	700 nM	15 nM
10 – 50 ng	251 – 500 ng	350 nM	10 nM
	100 – 250 ng	140 nM	10 nM

Additional Notes on Adapter Concentration

- Adapter concentration affects ligation efficiency as well as adapter and adapter-dimer carry-over in post-ligation cleanups. A molar excess of adapter is required to ensure optimal ligation efficiency. Low adapter:insert molar ratios (approaching 2:1) result in a significant proportion of insert molecules with an adapter ligated to only one end, leading to library construction failure.
- Adapter:insert molar ratios in the range of 10:1 – 40:1 are recommended for KAPA HTP and LTP Library Preparation Kits, whereas KAPA HyperPrep, KAPA HyperPlus Kits and KAPA RNA HyperPrep Kits are compatible with much higher ratios ($\geq 100:1$). Very high adapter:insert molar ratios (200:1 – 1000:1) may be beneficial for low-input library construction with KAPA HyperPrep and KAPA HyperPlus Kits.
- While it is not necessary to adjust adapter concentrations to accommodate moderate sample-to-sample variations, an adapter concentration that is appropriate for the range of input DNA concentrations is highly recommended.
- The best way to accommodate different adapter concentrations within a batch of samples processed together is to vary the concentration of adapter stock solutions and dispense a fixed volume (e.g., 5 μ L) of each adapter. The alternative—using a single stock solution and dispensing variable volumes of adapter into ligation reactions—is not recommended and is not compatible with higher throughput or automated workflows.
- Post-ligation cleanup and size selection strategies should be informed by the choice of adapter concentration. Please refer to **Important Parameters: Post-ligation Processing** for more details.
- Ultimately, the optimal adapter concentration for a specific workflow represents a compromise between ligation efficiency, the potential negative impact of adapter/adapter-dimer carry-over, and cost.

Adapter Concentration Calculations

- The below calculation applies to DNA library construction. For RNA library construction, adapter stock concentrations are calculated based on input only.
- To calculate the optimal adapter stock concentration for DNA library construction, the amount of input DNA (in picomoles) must first be calculated. This is done with the following formula:

$$\text{Picomoles} = \frac{\text{mass of DNA (ng)}}{660} \times \frac{1000}{\text{median size (bp)}}$$

- Next, the picomole quantity of adapter required is calculated by multiplying the number of picomoles of input DNA by the desired adapter:insert ratio. Please refer to Tables 2 and 3 or the Technical Data Sheet included with the library preparation kit for optimal adapter:insert molar ratios for different applications.
 - The picomole quantity of adapter required is subsequently divided by the volume of adapter used per reaction, to obtain the desired adapter stock concentration (in μ M or picomoles/ μ L).
- For example, 200 ng of input DNA with a mode fragment size of 250 bp represents 1.21 picomoles of insert DNA. For a 10:1 adapter:insert ratio, 12.1 picomoles of adapter is required. Therefore, when using 5 μ L of adapter stock per ligation reaction, an adapter stock concentration of 2.42 μ M is required.
- To obtain a calculator designed for the calculation of adapter:insert molar ratios and stock concentrations please contact Technical Support at sequencing.roche.com/support.

Post-ligation Processing

- It is important to remove excess unligated adapter and adapter-dimer molecules from Illumina libraries prior to library amplification or cluster generation. This is particularly important for libraries to be sequenced on Illumina platforms that employ patterned flow cells.
- KAPA HyperPure Beads or KAPA Pure Beads (jointly referred to hereafter as KAPA cleanup beads) or Agencourt® AMPure® XP (Beckman Coulter) are recommended for this application and are sold separately.
- Please follow the post-ligation cleanup instructions provided in the Technical Data Sheet for the library preparation kit. While a single post-ligation cleanup with KAPA cleanup beads or AMPure XP removes most unligated adapter and adapter-dimer (as recommended for KAPA HyperPrep and KAPA HyperPlus protocols), a second cleanup or size selection step may be necessary to eliminate any remaining adapter species from the library. The amount of adapter and adapter-dimer carried through the first cleanup is dependent on the library construction chemistry and adapter concentration in the ligation reaction.
- If bead-based size selection is carried out after adapter ligation, a single post-ligation cleanup (with the appropriate bead-to-sample ratio; as per the library construction protocol) must first be performed. Ligation buffers contain high concentrations of PEG 6000, which will impact the length and distribution of library fragments recovered from post-ligation size selection.

KAPA Dual-Indexed Adapter Kit

Illumina® Platforms

Technical Data Sheet

Index Sequences and Pooling Guidelines

- Sequencing indexes (barcodes) included in KAPA Dual-Indexed Adapters are provided in Table 10.
- For low-plexity pooling applications (up to 16-plex) on Illumina sequencing platforms, specific index combinations must be used. For 17- to 96-plex pools, any combination of adapters can be used. Detailed multiplexing guidelines are provided in Table 9.
- To ensure equal read distributions in multiplexed sequencing applications, libraries must be carefully quantified and/or normalized prior to pooling for capture or cluster generation. qPCR-based quantification with the KAPA Library Quantification Kit constitutes the most accurate and reproducible method for the quantification of sequenceable molecules in an Illumina library, particularly for PCR-free workflows.

Plate Format

Table 8 depicts the naming and placement of KAPA Dual-Indexed Adapters within the 96-well plate that is provided. Dual-indexed adapters are supplied in fully-skirted PCR plates, sealed with non-pierceable heat seals. Adapters are provided at a concentration of 15 µM. Each well contains 20 µL plus excess.

Table 8. KAPA Dual-Indexed Adapter Plate layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	A1	A2	A3	A4	A5	A6	A7	A8	A9	A10	A11	A12
B	B1	B2	B3	B4	B5	B6	B7	B8	B9	B10	B11	B12
C	C1	C2	C3	C4	C5	C6	C7	C8	C9	C10	C11	C12
D	D1	D2	D3	D4	D5	D6	D7	D8	D9	D10	D11	D12
E	E1	E2	E3	E4	E5	E6	E7	E8	E9	E10	E11	E12
F	F1	F2	F3	F4	F5	F6	F7	F8	F9	F10	F11	F12
G	G1	G2	G3	G4	G5	G6	G7	G8	G9	G10	G11	G12
H	H1	H2	H3	H4	H5	H6	H7	H8	H9	H10	H11	H12

Table 9. Detailed multiplexing guidelines

Plexity	Option	KAPA Dual-Indexed Adapters			
6	1	A1 – A3 and B1 – B3	C1 – C3 and D1 – D3	E1 – E3 and F1 – F3	G1 – G3 and H1 – H3
	2	A4 – A6 and B4 – B6	C4 – C6 and D4 – D6	E4 – E6 and F4 – F6	G4 – G6 and H4 – H6
	3	A7 – A9 and B7 – B9	C7 – C9 and D7 – D9	E7 – E9 and F7 – F9	G7 – G9 and H7 – H9
	4	A10 – A12 and B10 – B12	C10 – C12 and D10 – D12	E10 – E12 and F10 – F12	G10 – G12 and H10 – H12
8	1	A1 – A4 and B1 – B4	C1 – C4 and D1 – D4	E1 – E4 and F1 – F4	G1 – G4 and H1 – H4
	2	A5 – A8 and B5 – B8	C5 – C8 and D5 – D8	E5 – E8 and F5 – F8	G5 – G8 and H5 – H8
	3	A9 – A12 and B9 – B12	C9 – C12 and D9 – D12	E9 – E12 and F9 – F12	G9 – G12 and H9 – H12
12	1	A1 – A6 and B1 – B6	C1 – C6 and D1 – D6	E1 – E6 and F1 – F6	G1 – G6 and H1 – H6
	2	A7 – A12 and B7 – B12	C7 – C12 and D7 – D12	E7 – E12 and F7 – F12	G7 – G12 and H7 – H12
16	1	A1 – A8 and B1 – B8	C1 – C8 and D1 – D8	E1 – E8 and F1 – F8	G1 – G8 and H1 – H8
	2	A4 – A12 and B4 – B12	C4 – C12 and D4 – D12	E4 – E12 and F4 – F12	G4 – G12 and H4 – H12
>16		Any combination of the 96 available adapters			

Table 10. KAPA Dual-Indexed Adapter Kit Index sequences

Dual-Index	P5 Index	P7 Index
A1	TATAGCCT	ATTACTCG
A2	TATAGCCT	TCCGGAGA
A3	TATAGCCT	CGCTCATT
A4	TATAGCCT	GAGATTCC
A5	TATAGCCT	ATTCAGAA
A6	TATAGCCT	GAATTCGT
A7	TATAGCCT	CTGAAGCT
A8	TATAGCCT	TAATGCGC
A9	TATAGCCT	CGGCTATG
A10	TATAGCCT	TCCGCGAA
A11	TATAGCCT	TCTCGCGC
A12	TATAGCCT	AGCGATAG
B1	ATAGAGGC	ATTACTCG
B2	ATAGAGGC	TCCGGAGA
B3	ATAGAGGC	CGCTCATT
B4	ATAGAGGC	GAGATTCC
B5	ATAGAGGC	ATTCAGAA
B6	ATAGAGGC	GAATTCGT
B7	ATAGAGGC	CTGAAGCT
B8	ATAGAGGC	TAATGCGC
B9	ATAGAGGC	CGGCTATG
B10	ATAGAGGC	TCCGCGAA
B11	ATAGAGGC	TCTCGCGC
B12	ATAGAGGC	AGCGATAG
C1	CCTATCCT	ATTACTCG
C2	CCTATCCT	TCCGGAGA
C3	CCTATCCT	CGCTCATT
C4	CCTATCCT	GAGATTCC
C5	CCTATCCT	ATTCAGAA
C6	CCTATCCT	GAATTCGT
C7	CCTATCCT	CTGAAGCT
C8	CCTATCCT	TAATGCGC
C9	CCTATCCT	CGGCTATG
C10	CCTATCCT	TCCGCGAA
C11	CCTATCCT	TCTCGCGC
C12	CCTATCCT	AGCGATAG
D1	GGCTCTGA	ATTACTCG
D2	GGCTCTGA	TCCGGAGA
D3	GGCTCTGA	CGCTCATT
D4	GGCTCTGA	GAGATTCC
D5	GGCTCTGA	ATTCAGAA
D6	GGCTCTGA	GAATTCGT
D7	GGCTCTGA	CTGAAGCT
D8	GGCTCTGA	TAATGCGC
D9	GGCTCTGA	CGGCTATG
D10	GGCTCTGA	TCCGCGAA
D11	GGCTCTGA	TCTCGCGC
D12	GGCTCTGA	AGCGATAG

Dual-Index	P5 Index	P7 Index
E1	AGGCGAAG	ATTACTCG
E2	AGGCGAAG	TCCGGAGA
E3	AGGCGAAG	CGCTCATT
E4	AGGCGAAG	GAGATTCC
E5	AGGCGAAG	ATTCAGAA
E6	AGGCGAAG	GAATTCGT
E7	AGGCGAAG	CTGAAGCT
E8	AGGCGAAG	TAATGCGC
E9	AGGCGAAG	CGGCTATG
E10	AGGCGAAG	TCCGCGAA
E11	AGGCGAAG	TCTCGCGC
E12	AGGCGAAG	AGCGATAG
F1	TAATCTTA	ATTACTCG
F2	TAATCTTA	TCCGGAGA
F3	TAATCTTA	CGCTCATT
F4	TAATCTTA	GAGATTCC
F5	TAATCTTA	ATTCAGAA
F6	TAATCTTA	GAATTCGT
F7	TAATCTTA	CTGAAGCT
F8	TAATCTTA	TAATGCGC
F9	TAATCTTA	CGGCTATG
F10	TAATCTTA	TCCGCGAA
F11	TAATCTTA	TCTCGCGC
F12	TAATCTTA	AGCGATAG
G1	CAGGACGT	ATTACTCG
G2	CAGGACGT	TCCGGAGA
G3	CAGGACGT	CGCTCATT
G4	CAGGACGT	GAGATTCC
G5	CAGGACGT	ATTCAGAA
G6	CAGGACGT	GAATTCGT
G7	CAGGACGT	CTGAAGCT
G8	CAGGACGT	TAATGCGC
G9	CAGGACGT	CGGCTATG
G10	CAGGACGT	TCCGCGAA
G11	CAGGACGT	TCTCGCGC
G12	CAGGACGT	AGCGATAG
H1	GTACTGAC	ATTACTCG
H2	GTACTGAC	TCCGGAGA
H3	GTACTGAC	CGCTCATT
H4	GTACTGAC	GAGATTCC
H5	GTACTGAC	ATTCAGAA
H6	GTACTGAC	GAATTCGT
H7	GTACTGAC	CTGAAGCT
H8	GTACTGAC	TAATGCGC
H9	GTACTGAC	CGGCTATG
H10	GTACTGAC	TCCGCGAA
H11	GTACTGAC	TCTCGCGC
H12	GTACTGAC	AGCGATAG

Restrictions and Liabilities

This technical data sheet is provided “as is” and Kapa Biosystems assumes no responsibility for any typographical, technical, or other inaccuracies. The document is subject to change, without notice, in future editions.

To the maximum extent permitted by applicable law, Kapa Biosystems disclaims all warranties, either express or implied, with regard to this technical data sheet and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Kapa Biosystems shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein.

This document might contain references to third party sources of information, hardware or software, products, or services and/or third party web sites (collectively the “Third-Party Information”). Kapa Biosystems does not control, and is not responsible for, any Third-Party Information. The inclusion of Third-Party Information in this document does not imply endorsement by Kapa Biosystems of the Third-Party Information or the third party in any way.

Kapa Biosystems is not responsible nor will be liable in any way for your use of any software or equipment that is not supplied by Kapa Biosystems in connection with your use of Kapa Biosystems products.

Kapa Biosystems does not in any way guarantee or represent that you will obtain satisfactory results from using Kapa Biosystems products as described herein. The

only warranties provided to you are included in the Limited Warranty enclosed with this document. You assume all risk in connection with your use of Kapa Biosystems products.

Note to Purchaser: Limited Product Warranty

Any product that does not meet the performance standards stated in the product specification sheet will be replaced at no charge. This warranty limits our liability to the replacement of the product. No other warranties of any kind, express or implied, including without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Kapa Biosystems. Kapa Biosystems shall have no liability for any direct, indirect, consequential or incidental damages arising out of the use, the results of use or the inability to use any product.

Note to Purchaser: Limited License

KAPA Dual-Indexed Adapter Kits are developed, designed and sold exclusively for research purposes. Neither the product, nor any individual component, has been tested for use in diagnostics or for drug development, nor is it suitable for administration to humans or animals. Please refer to the SDS, which is available on request. Certain applications of this product are covered by patents issued to parties other than Kapa Biosystems and applicable in certain countries. Purchase of this product does not include a license to perform any such applications. Users of this product may therefore be required to obtain a patent license depending upon the particular application and country in which the product is used.



Manufacturing, R & D
Cape Town, South Africa
Tel: +27 21 448 8200
Fax: +27 21 448 6503

Technical Support
sequencing.roche.com/support

© 2019 KAPA and SEQCAP are trademarks of Roche. All other product names and trademarks are the property of their respective owners.